



THE SPECIALIST IN
GASES

CERTIFICATE OF CONFORMITY

<u>Certificate Number:</u> BGNL-1-204941	<u>Customer:</u> RaDu – Handelsond. B.V.
<u>Certificate Date</u> : 02-07-2025	Zeemanstraat 13
<u>Issued by</u> : BUSE Gas B.V.	2991 – XR Barendrecht
De Overmaat 21	
6831AE ARNHEM	
The Netherlands	
<u>Date of Analysis</u> : 02-07-2025	<u>Customer reference:</u> 2025500708
<u>Gas expiry Date</u> : 02-07-2027	<u>Ordering date:</u> 02-06-2025

Component:

Nominal Value:

Conformity Range:

Hydrogen Sulphide*
Nitrogen

25ppm
Balance

+/- 10%

*) H2S value is traceable to international standards based on analytical comparison to Dutch Metrology Institute, (VSL-NMI) Calibrated Gas Mixture: certificate no.C2544501 – cylinder no. D387455.

Cylinder Information:

Disposable type : 110L
Cylinder Size : 1,6 Litres (water volume)
Cylinder Fill Pressure/content : 68 bar, 110L gas
Valve Outlet Connection : C10
Storage/utilization temperature : -5°C to 50°C

Batch: N853-B/2625

Batch Size: 3 cylinder(s)

Handelsonderneming RaDu B.V.

Zeemanstraat 11a-13

2991 XR BARENDRECHT

Netherlands

Tel: +31 10 2921510

VAT nr. NL865121187B01

CoC 89816943

E-mail: info@radu.nl

Approved Signatory

Name:

Signature:

Date: 02-07-2025

